

Date: 16th September 2022

Product type	Brand / style	Manufacturer	ARTG	Relevant Standards	Comments
Bed	Alivio Electric	K Care	391217	EN 60335-1:2012 + A11:2014 EN 62233:2008 EN 55014-1:2006 + A1:2009 + A2:2011 EN 55014-2:1997 + A1:2001 + A2:2008	The applicable standards for this bed are related only to the actuators.
Mattress	Alivio	K Care	N/A	None Identified.	This product has been assessed as safe and fit for its intended purpose by the manufacturer.
Bed Accessory	Bed Sticks and Rails	K Care	391219	None Identified.	This product has been assessed as safe and fit for its intended purpose by the manufacturer.
Chairs and Seating	Fusion, Murray Bridge, Kingston, Hillite	K Care	391850	AS/NZS 4688.2:2000 AS/NZS 4688.3:2000	
	Alivio		216484	None relevant, see manufacturer Declaration of Conformity	
	Air Comfort V2, Maxi	K Care	374053	AS/NZS 3973:2009 (R2016) (Relevant Clauses)	
	Air Comfort Compact Lift Chairs	K Care	374053	AS/NZS 4688.4:2000	
	Princess Bed	K Care	391850	AS/NZS 3695.1:2011 AS/NZS ISO 7176.1:2015	
	Kingston Mobile	K Care	391850	AS/NZS 3973:2009 (R2016) (Relevant Clauses)	
	Evolution		391850	AS/NZS 4688.2:2000 (Relevant Clauses)	
Shower and bath chairs and stools	All Types	K Care	391391	AS/NZS 3973:2009 (R2016) AS/NZS 4688.2:2000, AS/NZS 4688.3:2000	
	K Care, Tubalco	K Care	392636	AS/NZS 3973:2009 (R2016)	
Commodes	Aquatec Ocean Dual VIP Ergo XL	Aquatec	153914	AS/NZS 3973:2009 AS/NZS ISO 7176.1:2015 AS/NZS ISO 7176.3:2015 AS/NZS ISO 7176.8:2015	
	Aquatec Ocean Ergo XL directional lock Mobile Shower Chair	Aquatec	153914	AS/NZS 3973:2009	
	Aquatec Ocean VIP Ergo Mobile Shower Chair	Aquatec	153914	AS/NZS 3973:2009	
	Aquatec Ocean Ergo Mobile Shower Chair with directional lock	Aquatec	153914	AS/NZS 3973:2009	
	Lift, patient transfer	Kerry	390530	AS ISO 10535-2002	
Lift, patient transfer slings	All Slings	K Care	390531	AS ISO 10535-2002	
Alternating Pressure Mattress Overlay System, chair cushion, Pressure alleviation pad, Negative pressure wound therapy pump	Entrix NX, Centrius, Dual Plus, Coziny Plus, Envelop, Theraflo AP, Airmax, Dual Flex, Cari Chair Overlays	Carilex Medical		Contact Carilex Medical ANZ for ARTG and Standards details. Email: nnati@kcare.com.au	
Walkers / Rollators	Wheeled walkers	K Care	390932	ISO 11199-2:2005 (Rollators)	
Walking Aid	Crutches and Walking frames with no wheels	K Care	390934	ISO 11199-2:1999	
Wheelchairs	Eagle KM-8000 (KM-8020/ KM-8520)	Karma Medical Products	182294	AS 3695:1992 AS 3696.1:2008 AS 3696-3:2008 AS 3696.5:1989 AS 3696.8:2009 ISO 7176-7:1996 ISO 7176-19:2008	
	Eagle HD (KM8020X/KM-8520X)	Karma Medical Products	182294	AS/NZS ISO 7176.1:2015 AS/NZS ISO 7176.3:2015 AS 3696.5:1989 ISO 7176.7:1998(E) AS/NZS ISO 7176.11:2013 AS 3696.19:2009 AS/NZS ISO 7176.22:2015 AS/NZS ISO 7176.26:2011	
	S Ergo	Karma Medical Products	182294	AS 3695:1992 AS 3696.1:2008 AS 3696-3:2008 AS 3696.5:1989 AS 3696.8:2009 ISO 7176-7:1996 ISO 7176-19:2008	
	Flexx, Flexx TL, Ergo Lite, VIP 2	Karma Medical Products	182294	AS 3695.1:2011 AS 3696.1:2008 AS 3696-3:2008 AS 3696.5:1989 AS 3696.8:2009 AS 3696.19:2009 AS 3696.22:2008 ISO 7176-5, 7, 11 & 16:1996 ISO 7176-19:2008	
	Flexx HD (KM-8022HD/KM-8522HD)	Karma Medical Products	182294	AS 3695.1:2011 AS 3696.1:2008 AS 3696-3:2008 AS 3696.5:1989 AS 3696.8:2009 AS 3696.19:2009 AS 3696.22:2008 ISO 7176-5, 7, 11 & 16:1996	
	VIP 515	Karma Medical Products	182294	AS 3695.1:2011 AS 3696.1:2008 AS 3696-3:2008 AS 3696.5:1989 AS 3696.8:2009 AS 3696.19:2009 AS 3696.22:2008 ISO 7176-5, 7, 11 & 16:1996 ISO 7176-19:2008 (with Tie Down Kit Fitted)	
	Ergo Lite 2	Karma Medical Products	182294	AS 3695.1:2011 AS 3696.1:2008 AS 3696-3:2008 AS 3696.5:1989 AS 3696.8:2009 AS 3696.19:2009 AS 3696.22:2008 ISO 7176-5, 7, 11 & 16:1996	
	Wheelchair, Electric	Karma Medical Products	182310	TBD	
One Touch Openers			N/A	No Relevant Standard Identified	This product has been assessed as safe and fit for its intended purpose by the manufacturer.
Baby Bath, electric Hi Lo.	FN20-860			No General Standard Applicable to this product Electric Components Compliant with: EN 60335-1 IEC60601-1 IEC 60601-6	This product has been assessed as safe and fit for its intended purpose by the manufacturer.
Knee Walker		K Care	381030	ISO 11199-2:2005 (R2016) Part 2: Rollators	New product
Walking Canes	Swan neck cane, T handle cane, Folding cane, Swan neck X Base cane	K Care	378630	BS EN 1985:1999 AS EN 12182:2015 AS/NZS ISO 11334.4:1999 (R2016)	New product